NITRAS MEDICAL

// ENSURE**SAFETY**

8310 // WHITE WAVE

Nitrile disposable gloves

PPE risk category III | Medical device class I

Very good tactile sensitivity

Very good dry and oil grip

Very robust and resistant

NI | AQL 1,5 | POWDER FREE | TESTED ACCORDING TO EN 455

VERSION 1.2.27 | 260927

8310 // WHITE WAVE

Material:	Nitrile, white
Coating:	-
Size(s):	XS-XL
Packaging unit:	10 piece(s)
Certification:	EN ISO 21420, EN ISO 374, EN 421

NITRAS WHITE WAVE, nitrile disposable gloves, white (colour code: 1100), non-sterile, rolled edge, powder free, manufactured according to EN 455, for food contact, ambidextrous (left and right hand), medical gloves, micro-roughened fingertips, colour code system for sizes, AQL 1.5, protection against micro-organisms, bacteria and viruses, protection against radioactive contamination by solid particles, particularly sensitive, perfect for working with small parts, very good tactile sensitivity, very good dry and oil grip, very robust and resistant, medical devices class I, PPE risk category III, box per 100 pieces

ISO 374-1/Type B JKT | ISO 374-5:2016 VIRUS | EN ISO 21420 | EN 421:2010

NI | AQL 1,5 | POWDER FREE | TESTED ACCORDING TO EN 455

Industries

- Beauty Edition
- Care Edition
- Cleaning Edition
- Dental Edition
- Food Edition
- Medical Edition

Advantages

+ Very good tactile sensitivity
+ Very good dry and oil grip
+ Very robust and resistant

Further images

Colour variants

8310 // WHITE WAVE
Colour code: 1100

8324 // INDIGO WAVE
Colour code: 2400

8328 // VIOLET WAVE
Colour code: 6200

8327 // CANDY WAVE
Colour code: 6400

8326 // RED WAVE
Colour code: 6000

8325 // MINT WAVE
Colour code: 3200

8323 // GREEN WAVE
Colour code: 3000

8322 // PURPLE WAVE
Colour code: 6200

8321 // PINK WAVE
Colour code: 6300

8320 // BLACK WAVE
Colour code: 1000

8311 // BLUE WAVE
Colour code: 2200

8304 // GLAMOUR WAVE
Colour code: 6300

Performance values

EN ISO 21420:2020 + A1:2024	Test parameter	Performance levels	Test result
	Dexterity	1-5	5

EN ISO 374-1:2016 + A1:2018	Test parameter	Performance levels	Test result
JKT	J - n-Heptane	1-6	3 (> 60)
	K - Sodium hydroxide 40%	1-6	6 (> 480)
	T - Formaldehyde 37%	1-6	4 (> 120)

EN ISO 374-5:2016	Test parameter	Performance levels	Test result
VIRUS	Penetration	-	Passed
	Protection against viruses	-	Passed

EN 421:2010	Test parameter	Performance levels	Test result

Additional information

	Material	Colour
Palm	Nitrile	white
Index finger and thumb	Nitrile	white
Back of the hand	Nitrile	white
Fingertips	Nitrile	white
Cuff / knitted wrist	Nitrile	white

Thickness (approx.)	
Finger	0,10 mm
Palm	0,08 mm
Cuff	0,07 mm

Size	Length (mm)	Width palm (mm)
XS	240	75
S	240	85
M	240	95
L	240	105
XL	240	115

Properties	Before ageing	After ageing
Elongation at break	600%	500%
Tear resistance	18 MPa	20 MPa

Logistics data

Size	GTIN (SPU)	GTIN (PU)	Length (m)	Width (m)	Height (m)	Weight (kg)	Quantity PU / pallet
XS	4059904035430	4059904003842	0,34 m	0,24 m	0,24 m	3,73 kg	80
S	4059904035447	4059904003859	0,34 m	0,24 m	0,24 m	3,98 kg	80
M	4059904035454	4059904003866	0,34 m	0,24 m	0,24 m	4,3 kg	80
L	4059904035461	4059904003873	0,34 m	0,24 m	0,24 m	4,44 kg	80
XL	4059904035478	4059904003880	0,34 m	0,24 m	0,24 m	5,23 kg	80

Customs tariff number: 40151200000
Country of origin: Malaysia

MCD Medical Care Dental GmbH | Heinrich-Hertz-Str. 7 | 50181 Bedburg | Germany
+49 2272 401 111 info@nitras-medical.com www.nitras-medical.com
NITRAS MEDICAL nitras_medical NITRAS MEDICAL NITRAS MEDICAL

Alle angegebenen Informationen entsprechen dem Stand vom angegebenen Datum. Für Rückfragen und aktuelle Auskünfte stehen wir Ihnen jederzeit gerne zur Verfügung. Alle Angaben ohne Gewähr.