# 8320 // BLACK WAVE

Nitrile disposable gloves

PPE risk category III | Medical device class I




Very good tactile sensitivity

Very good dry and oil grip

Very robust and resistant

AQL 1,5 | NI | TESTED ACCORDING TO EN 455

VERSION 1.3.11 | 261105

# 8320 // BLACK WAVE

| | |
|---|---|
| **Material:** | Nitrile, black |
| **Coating:** | - |
| **Size(s):** | XS-XL |
| **Packaging unit:** | 10 piece(s) |
| **Certification:** | EN ISO 21420, EN ISO 374, EN 421 |

NITRAS BLACK WAVE, nitrile disposable gloves, black (colour code: 1000), non-sterile, rolled edge, powder free, manufactured according to EN 455, for food contact, ambidextrous (left and right hand), medical gloves, micro-roughened fingertips, colour code system for sizes, AQL 1.5, protection against micro-organisms, bacteria and viruses, protection against radioactive contamination by solid particles, particularly sensitive, perfect for working with small parts, very good tactile sensitivity, very good dry and oil grip, very robust and resistant, medical devices class I, PPE risk category III, box per 100 pieces

ISO 374-1:2016/Type B JKT

ISO 374-5:2016 VIRUS

EN ISO 21420

EN 421:2010

AQL 1,5

NI

TESTED ACCORDING TO EN 455

## Industries

- Beauty Edition
- Black Edition
- Cleaning Edition
- Dental Edition
- Food Edition
- Medical Edition

## Advantages

- \+ Very good tactile sensitivity
- \+ Very good dry and oil grip
- \+ Very robust and resistant



**Further images**





## Colour variants




**8320 // BLACK WAVE**
Colour code: 1000




**8304 // GLAMOUR WAVE**
Colour code: 6300




**8311 // BLUE WAVE**
Colour code: 2200




**8310 // WHITE WAVE**
Colour code: 1100




**8321 // PINK WAVE**
Colour code: 6300




**8322 // PURPLE WAVE**
Colour code: 6200




**8323 // GREEN WAVE**
Colour code: 3000




**8324 // INDIGO WAVE**
Colour code: 2400




**8325 // MINT WAVE**
Colour code: 3200




**8326 // RED WAVE**
Colour code: 6000




**8327 // CANDY WAVE**
Colour code: 6400




**8328 // VIOLET WAVE**
Colour code: 6200

**Performance values**

| EN ISO 21420:2020 + A1:2024 | Test parameter | Performance levels | Test result |
|---|---|---|---|
| | Dexterity | 1-5 | 5 |

| EN ISO 374-1:2016 + A1:2018 | Test parameter | Performance levels | Test result |
|---|---|---|---|
|  JKT | J - n-Heptane | 1-6 | 3 (> 60) |
| | K - Sodium hydroxide 40% | 1-6 | 6 (> 480) |
| | T - Formaldehyde 37% | 1-6 | 4 (> 120) |

| EN ISO 374-5:2016 | Test parameter | Performance levels | Test result |
|---|---|---|---|
|  VIRUS | Penetration | - | Passed |
| | Protection against viruses | - | Passed |

| EN 421:2010 | Test parameter | Performance levels | Test result |
|---|---|---|---|
|  | | | |

## Additional information

| | Material | Colour |
|---|---|---|
| Palm | Nitrile | black |
| Index finger and thumb | Nitrile | black |
| Back of the hand | Nitrile | black |
| Fingertips | Nitrile | black |
| Cuff / knitted wrist | Nitrile | black |

| Thickness (approx.) | |
|---|---|
| Finger | 0,10 mm |
| Palm | 0,08 mm |
| Cuff | 0,07 mm |

| Size | Length (mm) | Width palm (mm) |
|---|---|---|
| XS | 240 | 76 |
| S | 240 | 84 |
| M | 240 | 94 |
| L | 240 | 105 |
| XL | 240 | 113 |

| Properties | Before ageing | After ageing |
|---|---|---|
| Elongation at break | 500% | 400% |
| Tear resistance | 14 MPa | 14 MPa |

## Logistics data

| Size | GTIN (SPU) | GTIN (PU) | Length (m) | Width (m) | Height (m) | Weight (kg) | Quantity PU / pallet |
|---|---|---|---|---|---|---|---|
| XS | 4059904035539 | 4059904003941 | 0,37 m | 0,24 m | 0,22 m | 4,15 kg | 80 |
| S | 4059904035546 | 4059904003958 | 0,37 m | 0,24 m | 0,22 m | 4,68 kg | 80 |
| M | 4059904035553 | 4059904003965 | 0,37 m | 0,24 m | 0,22 m | 4,55 kg | 80 |
| L | 4059904035560 | 4059904003972 | 0,37 m | 0,24 m | 0,22 m | 5,25 kg | 80 |
| XL | 4059904035577 | 4059904003989 | 0,37 m | 0,24 m | 0,22 m | 5,41 kg | 80 |

Customs tariff number: 40151200000
Country of origin: Malaysia



MCD Medical Care Dental GmbH | Heinrich-Hertz-Str. 7 | 50181 Bedburg | Germany

+49 2272 401 111 info@nitras-medical.com www.nitras-medical.com

NITRAS MEDICAL nitras_medical NITRAS MEDICAL NITRAS MEDICAL

Alle angegebenen Informationen entsprechen dem Stand vom angegebenen Datum. Für Rückfragen und aktuelle Auskünfte stehen wir Ihnen jederzeit gerne zur Verfügung. Alle Angaben ohne Gewähr.