NITRAS MEDICAL

// ENSURE**SAFETY**

8322 // PURPLE WAVE

Nitrile disposable gloves

PPE risk category III | Medical device class I

Very good tactile sensitivity

Very good dry and oil grip

Very robust and resistant

AQL 1,5 | NI | TESTED ACCORDING TO EN 455

VERSION 1.3.8 | 260927

8322 // PURPLE WAVE

Material:	Nitrile, purple
Coating:	-
Size(s):	XS-XL
Packaging unit:	10 piece(s)
Certification:	EN ISO 21420, EN ISO 374, EN 421

NITRAS PURPLE WAVE, nitrile disposable gloves, purple (colour code: 6200), non-sterile, rolled edge, powder free, manufactured according to EN 455, for food contact, ambidextrous (left and right hand), medical gloves, micro-roughened fingertips, colour code system for sizes, AQL 1.5, protection against micro-organisms, bacteria and viruses, protection against radioactive contamination by solid particles, particularly sensitive, perfect for working with small parts, very good tactile sensitivity, very good dry and oil grip, very robust and resistant, medical devices class I, PPE risk category III, box per 100 pieces

ISO 374-1:2016/Type B — JKT

ISO 374-5:2016 — VIRUS

EN ISO 21420

EN 421:2010

AQL 1,5

NI

TESTED ACCORDING TO EN 455

Industries

- Beauty Edition
- Black Edition
- Cleaning Edition
- Dental Edition
- Food Edition
- Medical Edition

Advantages

+ Very good tactile sensitivity
+ Very good dry and oil grip
+ Very robust and resistant

Further images

Colour variants

8322 // PURPLE WAVE
Colour code: 6200

8304 // GLAMOUR WAVE
Colour code: 6300

8310 // WHITE WAVE
Colour code: 1100

8311 // BLUE WAVE
Colour code: 2200

8320 // BLACK WAVE
Colour code: 1000

8321 // PINK WAVE
Colour code: 6300

8323 // GREEN WAVE
Colour code: 3000

8324 // INDIGO WAVE
Colour code: 2400

8325 // MINT WAVE
Colour code: 3200

8326 // RED WAVE
Colour code: 6000

8327 // CANDY WAVE
Colour code: 6400

8328 // VIOLET WAVE
Colour code: 6200

Performance values

EN ISO 21420:2020 + A1:2024	Test parameter	Performance levels	Test result
	Dexterity	1-5	5

EN ISO 374-1:2016 + A1:2018	Test parameter	Performance levels	Test result
JKT	J - n-Heptane	1-6	3 (> 60)
	K - Sodium hydroxide 40%	1-6	6 (> 480)
	T - Formaldehyde 37%	1-6	4 (> 120)

EN ISO 374-5:2016	Test parameter	Performance levels	Test result
VIRUS	Penetration	-	Passed
	Protection against viruses	-	Passed

EN 421:2010	Test parameter	Performance levels	Test result

Additional information

	Material	Colour
Palm	Nitrile	purple
Index finger and thumb	Nitrile	purple
Back of the hand	Nitrile	purple
Fingertips	Nitrile	purple
Cuff / knitted wrist	Nitrile	purple

Thickness (approx.)	
Finger	0,10 mm
Palm	0,08 mm
Cuff	0,07 mm

Size	Length (mm)	Width palm (mm)
XS	240	76
S	240	84
M	240	94
L	240	105
XL	240	113

Properties	Before ageing	After ageing
Elongation at break	500%	400%
Tear resistance	14 MPa	14 MPa

Logistics data

Size	GTIN (SPU)	GTIN (PU)	Length (m)	Width (m)	Height (m)	Weight (kg)	Quantity PU / pallet
XS	4059904035638	4059904004047	0,37 m	0,24 m	0,22 m	3,56 kg	80
S	4059904035645	4059904004054	0,37 m	0,24 m	0,22 m	4,03 kg	80
M	4059904035652	4059904004061	0,37 m	0,24 m	0,22 m	4,35 kg	80
L	4059904035669	4059904004078	0,37 m	0,24 m	0,22 m	4,44 kg	80
XL	4059904035676	4059904004085	0,37 m	0,24 m	0,22 m	4,85 kg	80

Customs tariff number: 40151200000
Country of origin: Malaysia

MCD Medical Care Dental GmbH | Heinrich-Hertz-Str. 7 | 50181 Bedburg | Germany
+49 2272 401 111 info@nitras-medical.com www.nitras-medical.com
NITRAS MEDICAL nitras_medical NITRAS MEDICAL NITRAS MEDICAL

Alle angegebenen Informationen entsprechen dem Stand vom angegebenen Datum. Für Rückfragen und aktuelle Auskünfte stehen wir Ihnen jederzeit gerne zur Verfügung. Alle Angaben ohne Gewähr.